

2020/09/29: Tuesday

ORMF, 43 MHz

off gas

of room temp. flux curves

X-ray

- Detector needed to be re-shielded

- Mylar permits Apertur 5 to be used!

- Real X-ray seen: Bremsstrahlung & O K α line

(confirmed by using opt.)