

Band code	Band type	Sampling rate (sec)	Corner period (sec)
F	...	> 0.0002 to ≤ 0.001	≥ 10 sec
G	...	> 0.0002 to ≤ 0.001	< 10 sec
D	...	> 0.001 to ≤ 0.004	< 10 sec
C	...	> 0.001 to ≤ 0.004	≥ 10 sec
E	Extremely Short Period	≤ 0.0125	< 10 sec
S	Short Period	≤ 0.1 to > 0.0125	< 10 sec
H	High Broad Band	≤ 0.0125	≥ 10 sec
B	Broad Band	≤ 0.1 to > 0.0125	≥ 10 sec
M	Mid Period	< 1 to > 0.1	
L	Long Period	1	
V	Very Long Period	10	
U	Ultra Long Period	100	
R	Extremely Long Period	1000	
P	On the order of 0.1 to 1 day	≤ 100000 to > 10000	
T	On the order of 1 to 10 days	≤ 1000000 to > 100000	
Q	Greater than 10 days	> 1000000	
A	Administrative Instrument Channel	variable	NA
O	Opaque Instrument Channel	variable	NA